APPENDIX B

Third Quarter 2010 IML Calibration Report

Energy Fuels Resources Corporation
Piñon Ridge Mill

Calibration Report for
Meteorological Monitoring Network

3rd Quarter 2010

Prepared by:

IML Air Science
a division of Inter-Mountain Laboratories, Inc.
555 Absaraka
Sheridan, Wyoming 82801
(307) 674-7506
www.imlairscience.com

1 Introduction

Inter-Mountain Laboratories – Air Science Division performed calibrations on July 27, 2010. The calibrations included the (2) meteorological monitoring systems at the Piñon Ridge Mill Site located approximately 15 miles from Naturita, Colorado. The Piñon Ridge Mill Site is operated by Energy Fuels Resources Corporation. This is a list of the monitoring sites and the associated equipment:

Site 1

- Meteorological Station – 10m Tower
  - Wind Speed
  - Wind Direction
  - Vertical Wind Speed
  - Temperature (2m & 10m)
  - Delta Temperature
  - Relative Humidity
  - Solar Radiation
  - Barometric Pressure
  - Precipitation
  - Evaporation

Site 2

- Meteorological Station – 30m Tower
  - Wind Speed
  - Wind Direction
  - Vertical Wind Speed
  - Temperature (2m & 30m)
  - Delta Temperature
  - Relative Humidity
  - Solar Radiation
  - Barometric Pressure

1.1 Calibration References

The calibrations were conducted in accordance with the following guideline documents:

- Ambient Monitoring Guidelines for the Prevention of Significant Deterioration (PSD), May 1987
- Environmental Protection Agency (EPA) Meteorological Monitoring Guidance for Regulatory Modeling Applications, February 2000 (MMGRMA) (EPA-454/R-99-005)
- Quality Assurance Handbook for Air Pollution Measurements Systems, Vol. I – A Field Guide to Environmental Quality Assurance, April 1994
- Quality Assurance Handbook for Air Pollution Measurements Systems, Vol. IV – Meteorological Measurements, March 2008
- Quality Assurance Handbook for Air Pollution Measurement Systems, Vol. V, Meteorological Measurements, EPA 1995
- Ambient Air Monitoring Requirements for the Air Pollution Control Division of the Colorado Department of Public Health and Environment, Technical Services Program Air Pollution Control Division, April 2001
- U.S. Nuclear Regulatory Commission Regulatory Guide, Office of Nuclear Regulatory Research, Regulatory Guide 3.63 – Onsite Meteorological Measurement Program For Uranium Recovery Facilities – Data Acquisition and Reporting, March 1988.

2 Calibration Methodology and Accuracy Goals

2.1 Meteorological Stations

2.1.1 Wind Speed

The wind speed was verified by rotating the sensor shaft using a DC-powered variable-speed motor equipped with an optical encoder output referenced to a crystal oscillator. A standard sensor speed was calculated based on the audit rotational speed and compared to the instantaneous logger reading. An R.M. Young Torque Disc was used to ensure bearing integrity of the wind speed sensor. All data were recorded on a standardized audit form.

2.1.2 Wind Direction

The wind direction sensor orientation was verified by using a Brunton precision magnetic compass. Instantaneous direction readings from the logger were compared to the standards and recorded on a standardized form.

2.1.3 Temperature

Proper operation of the temperature sensors was verified by placing the sensors and a precision NIST-traceable electronic thermometer in three equilibrated temperature baths (ice bath, warm bath, and ambient bath). Both reference thermometer and logger readings were recorded on a standardized form.

2.1.4 Delta Temperature

Proper operation of the temperature sensors was verified by placing the sensors and a precision NIST-traceable electronic thermometer in three equilibrated temperature baths (ice bath, warm bath, and ambient bath). Both reference thermometer and logger readings were recorded on a standardized form.

2.1.5 Relative Humidity

The relative humidity was checked by co-locating a reference sensor next to the station sensor. The reading was taken and the difference between the calibration standard and the on-site data logger were compared to acceptance criteria.

2.1.6 Solar Radiation

The solar radiation was checked by co-locating a reference sensor next to the station sensor. The readings of covered and uncovered were taken and the differences between the calibration standard and the on-site data logger were compared to acceptance criteria.

2.1.7 Barometric Pressure

The barometric pressure was checked by co-locating a reference sensor next to the station sensor. The reading was taken and the difference between the calibration standard and the on-site data logger were compared to acceptance criteria.

2.1.8 Precipitation

The precipitation gauge was challenged three times using a lab quality burette and water. The volume of water required to cause the tipping bucket to activate was measured and volumes were recorded along with the calculated value for activation on a standardized form.

2.1.9 Evaporation

The evaporation pan was calibrated by using five points to calculate the resulting slope and intercept for the sensor. The calibration is completed by adding water to the evaporation pan and recording the values from a yardstick and the logger.

2.2 Calibration Thresholds

Calibration goals for the parameters measured by the meteorological monitoring system are those specified in the US EPA *Quality Assurance Handbook for Air Pollution Measurement Systems, Volume IV, Meteorological Measurements*, March 2008. Accuracy goals by parameter are shown below.

Table 2-1 – Meteorological Sensor Criteria

Sensor	Specifications
Wind Speed	±0.5 m/s
Wind Speed – Starting Threshold	< 0.5 gm-cm
Wind Direction	± 5.0 compass degrees
Vertical Wind Speed	±.2 m/s ± 5.0 percent of observed
Temperature	±0.5 °C
Delta Temperature	±0.1 °C
Relative Humidity	±5.0 %
Solar Radiation	±5.0 percent of observed
Barometric Pressure	±0.09 in Hg
Precipitation	±10 percent of observed

3 Calibration Results

Calibration results for Site #1 – 2 can be found in Appendices A.

4 Findings/Recommendations

The calibrations of the Meteorological Monitoring equipment were within the calibration specifications at Site #1 and Site #2, including the evaporation pan. Results of the calibrations can be found in Appendix A.

Appendix A

Meteorological System Calibrations

METEOROLOGICAL STATION CALIBRATION SUMMARY

Met Station: Energy Fuels - Site #1 (10m Tower) Calibration Date: 27-Jul-10

Calibration Performed By: W. Adler, J. Masters -- IML Air Science

Sensor	Mfr./Model	Serial Number	Reference Device	Serial/ID Number
Wind Speed 10m (WS):	RM Young Wind Monitor AQ	82346	quartz referenced drive motor	IML 0896
Wind Direction 10m (WD):	RM Young Wind Monitor AQ	82346	transit, compass	IML 0942
Temperature @ 2 Meters:	RM Young Platinum RTD Temp Probe	13638	digital thermistor	IML 1401
Temperature @ 10 Meters:	RM Young Platinum RTD Temp Probe	13639	digital thermistor	IML 1401
Relative Humidity (RH):	CSI CS 500	C2430074	digital hygrometer	IML 0899
Barometric Pressure (BP):	PTB 101B	C2430048	digital barometer	IML 1404
Solar Radiation:	LI-COR LI200SZ	PY57101	collocated LI200X	PY54289
Precipitation:	Met One 12" tipping bucket	G6356	lab grade burette	N/A
Data acquisition system:	CSI CR3000 datalogger	2397	N/A	N/A

Calibration Results

	Reference RPM	Reference m/s	DAS Value	|Difference|	Specification	
AQ WS 10 meters (cm/s)	0	0.00	0.00	0.00	below threshold	
	300	1.54	1.54	0.00	0.20	(2)
	800	4.10	4.10	0.00	0.20	(2)
	3000	15.36	15.36	0.00	0.20	(2)
	8000	40.96	40.96	0.00	0.20	(2)
		Reference	DAS Value	|Difference|	Specification	
WS 10 m start torque (gm-cm)		τ<0.5	N/A	N/A	1.0	(3)
WD 10 meters (degrees)		0.0	0.8	0.8	5.0	(2)
		90.0	89.8	0.2	5.0	(2)
		180.0	180.5	0.5	5.0	(2)
		270.0	270.4	0.4	5.0	(2)

		Reference RPM	Reference cm/s	DAS Value	|Difference|	Specification	
Vertical WS 10 meters (cm/s)		0	0.00	-0.29	0.29	below threshold	
(Clockwise)		20	10.00	9.87	0.13	20.00	(2)
	EPS:	200	100.00	99.26	0.74	20.00	(2)
		300	150.00	147.73	2.27	20.00	(2)
		500	250.00	248.48	1.52	20.00	(2)
		0	0.00	-0.29	0.29	below threshold	
		20	10.00	9.78	0.23	20.00	(2)
	CFT:	20	100.00	99.89	0.11	20.00	(2)
		300	150.00	149.59	0.41	20.00	(2)
		500	250.00	248.53	1.47	20.00	(2)
Vertical WS 10 meters (cm/s)		0	0.00	-0.29	0.29	below threshold	
(Counter-Clockwise)		20	-10.00	-10.16	0.16	20.00	(2)
	EPS:	200	-100.00	-99.86	0.14	20.00	(2)
		300	-150.00	-146.31	3.69	20.00	(2)
		500	-250.00	-247.93	2.07	20.00	(2)
		0	0.00	-0.29	0.29	below threshold	
		20	-10.00	-10.36	0.36	20.00	(2)
	CFT:	200	-100.00	-99.53	0.47	20.00	(2)
		300	-150.00	-148.71	1.29	20.00	(2)
		500	-250.00	-247.94	2.06	20.00	(2)
WS 10 m start torque (gm-cm)			τ<0.5	N/A	N/A	1.0	(3)

		Reference (°C):	DAS Value	|Difference|	Specification	
		42.74	42.69	0.05	0.5	(2)
Temp. (°C): 2 meter		27.80	28.00	0.20	0.5	(2)
		2.82	3.16	0.34	0.5	(2)
Temp. (°C): 10 meter		42.74	42.70	0.04	0.5	(2)
		27.80	27.93	0.13	0.5	(2)
		2.82	3.13	0.31	0.5	(2)

		Reference	DAS Value	|Difference|	Specification	
Relative Humidity (%)		27.6	32.2	4.6	5.0	(2)
RH Sensor Temp (°C):		88.1	91.3	3.2		
Solar Radiation (W/m^2)	un-covered	1094	1063.00	31.0	54.7	(2)
	covered	0.00	0.00			
Barometric Pressure ("Hg)		24.60	24.64	0.04	0.09	(4)

	DAS Value (in)	Reference (ml)	DAS Equivalent	|Difference|	Specification	
Precipitation (0.1" equiv.)	0.10	175.5	185.3	9.8	18.5	(2)
	0.10	189.5	185.3	4.2	18.5	(2)
	0.10	187.7	185.3	2.4	18.5	(2)
			Average Diff:	5.5	18.5	(2)

	2m sensor	10m sensor	ΔT - B	Specification	
Delta Temperature (°C):	42.69	42.70	0.01	0.1	(2)
	28.00	27.93	0.07	0.1	(2)
	3.16	3.13	0.03	0.1	(2)

BOLD difference values exceed performance specifications

(1)= Performance specification listed in facilities' Quality Assurance Project Plan

(2)= EPA Quality Assurance Manual for Air Pollution Measurement Systems, Vol. IV, 2008

(3)= Manufacturer's Specifications

(4)= EPA On-Site Meteorological Program Guidance for Regulatory Modeling Applications

Notes, Recommendations

System taken offline at 12:33 MST and returned online at 13:19 MST.

METEOROLOGICAL STATION CALIBRATION SUMMARY

Met Station: Energy Fuels - Site #2 (30m Tower) Calibration Date: 27-Jul-10

Calibration Performed By: W. Adler, J. Masters - - IML Air Science

Sensor	Mfr./Model	Serial Number	Reference Device	Serial/ID Number
Wind Speed 30m (WS):	RM Young Wind Monitor AQ	82347	quartz referenced drive motor	IML 0896
Wind Direction 30m (WD):	RM Young Wind Monitor AQ	82347	transit, compass	IML 0942
Temperature @ 2 Meters:	RM Young Platinum RTD Temp Probe	13640	digital thermistor	IML 1401
Temperature @ 30 Meters:	RM Young Platinum RTD Temp Probe	13641	digital thermistor	IML 1401
Relative Humidity (RH):	CSI CS 500	C2730148	digital hygrometer	IML 0899
Barometric Pressure (BP):	PTB 101B	C2750056	digital barometer	IML 1404
Solar Radiation:	LI-COR LI200SZ	PY57102	collocated LI200X	PY54289
Data acquisition system:	CSI CR3000 datalogger	2421	N/A	N/A

Calibration Results

		Reference RPM	Reference m/s	DAS Value	|Difference|	Specification	
AQ WS 30 meters (cm/s)		0	0.00	0.00	0.00	below threshold	
		300	1.54	1.54	0.00	0.20	(2)
		800	4.10	4.10	0.00	0.20	(2)
		3000	15.36	15.36	0.00	0.20	(2)
		8000	40.96	40.96	0.00	0.20	(2)
			Reference	DAS Value	|Difference|	Specification	
WS 30 m start torque (gm-cm)			τ<0.5	N/A	N/A	1.0	(3)
WD 30 meters (degrees)			0.0	0.2	0.2	5.0	(2)
			90.0	90.3	0.3	5.0	(2)
			180.0	180.9	0.9	5.0	(2)
			270.0	270.6	0.6	5.0	(2)

		Reference RPM	Reference cm/s	DAS Value	|Difference|	Specification	
Vertical WS 30 meters (cm/s)		0	0.00	0.00	0.00	below threshold	
(Clockwise)		20	10.00	9.66	0.34	20.00	(2)
	EPS:	200	100.00	98.40	1.60	20.00	(2)
		300	150.00	149.36	0.64	20.00	(2)
		500	250.00	246.31	3.69	20.00	(2)
		0	0.00	-0.29	0.29	below threshold	
		20	10.00	10.16	0.16	20.00	(2)
	CFT:	200	100.00	99.51	0.49	20.00	(2)
		300	150.00	146.04	3.96	20.00	(2)
		500	250.00	249.63	0.37	20.00	(2)
Vertical WS 30 meters (cm/s)		0	0.00	0.00	0.00	below threshold	
(Counter-Clockwise)		20	-10.00	-9.90	0.10	20.00	(2)
	EPS:	200	-100.00	-100.46	0.46	20.00	(2)
		300	-150.00	-151.71	1.71	20.00	(2)
		500	-250.00	-249.53	0.47	20.00	(2)
		0	0.00	0.00	0.00	below threshold	
		20	-10.00	-9.86	0.14	20.00	(2)
	CFT:	200	-100.00	-100.41	0.41	20.00	(2)
		300	-150.00	-147.03	2.97	20.00	(2)
		500	-250.00	-248.35	1.65	20.00	(2)
WS 30 m start torque (gm-cm)			τ<0.5	N/A	N/A	1.0	(3)

		Reference (°C):	DAS Value	|Difference|	Specification	
		7.12	7.25	0.13	0.5	(2)
Temp. (°C): 2 meter		27.40	27.85	0.45	0.5	(2)
		40.52	40.57	0.05	0.5	(2)
Temp. (°C): 30 meter		7.12	7.19	0.07	0.5	(2)
		27.40	27.84	0.44	0.5	(2)
		40.52	40.53	0.01	0.5	(2)

		Reference	DAS Value	|Difference|	Specification	
Relative Humidity (%)		30.3	29.8	0.5	7.0	(2)
RH Sensor Temp (°C):		85.7	85.7	0.0		
Solar Radiation (W/m^2)	un-covered	110.81	107.56	3.3	5.5	(2)
	covered	0.00	0.63			
Barometric Pressure ("Hg)		24.55	24.55	0.00	0.09	(4)

	2m sensor	10m sensor	ΔT - B	Specification	
Delta Temperature (°C):	7.25	7.19	0.06	0.1	(2)
	27.85	27.84	0.01	0.1	(2)
	40.57	40.53	0.04	0.1	(2)

BOLD difference values exceed performance specifications

(1)= Performance specification listed in facilities' Quality Assurance Project Plan

(2)= EPA Quality Assurance Manual for Air Pollution Measurement Systems, Vol. IV, 2008

(3)= Manufacturer's Specifications

(4)= EPA On-Site Meteorological Program Guidance for Regulatory Modeling Applications

Notes, Recommendations

System taken offline at 12:45 MST and returned online at 17:33 MST.

Evaporation Pan Verification Sheet

Evap Span	**9.7221**
Evap Offset	**1.48158**

Yardstick Reading (in)*	Logger Reading	Percent Difference
6.25	**6.47**	**3.5%**

Notes:
* Yard Stick Reading - taken at the pan outlet to the gauge
Offline: 1233 MST
Online: 1319 MST

Appendix B

Transfer Standard Certifications

Certificate of Accuracy

Transfer Standard Type: Electronic Thermometer Certificate No: T **020910. 03**

Transfer standard, model/type: Lufft Model C100 Thermometer
Serial number: 024.0105.0202.4.2.1.19 **IML 1401**
submitted by/owner: Inter-Mountain Laboratories, Inc. Probe S/N: 21631.041
Air Science Division
555 Absaraka Street
Sheridan, WY 82801

Was compared to:
NIST Traceable Hg-in-glass thermometers, serial numbers 2J3106, 2Y6027, 3L9452.
Miller & Weber Hg-in-glass thermometer sn 2J3106 and 2Y6027 are traceable to NIST Test No. 209621, Test Method ASTM E-77. 2J3106 is traceable through Standard No. 1S1262. 2Y6027 is traceable through Standard No. 9C8072. Miller & Weber Hg-in-glass thermometer sn 3L9452 is traceable to NIST Thermometer 40350, through Transfer Standards 3C4465 & 1Y9716.

Date: 02/09/10 Lab temperature: 73.5 °F
OFFSET = 0.10 Barometric Pressure: 659.6 mmHG

Reference Standard (°C)	Transfer Standard (°C)	Difference from Reference (°C)	Transfer Standard Correction* (°C)
0.0	0.00	0.0	0.0
10.0	10.00	0.0	0.0
17.9	17.90	0.0	0.0
24.0	24.00	0.0	0.0
31.5	31.50	0.0	0.0

OFFSET = 0.10

Note:
If no sign is given on the correction, the true temperature is higher than the indicated temperature. If the sign is negative, the true temperature is lower than the indicated temperature.

Transfer Standard adjustments made? **YES** ☒ NO ☐
Post-calibration measurements:

Reference Thermistor (°C)	Transfer Standard (°C)	Difference from Reference (°C)	Transfer Standard Correction* (°C)
See Offset and measurements above			

Certified By: R W [signature] Date: February 9, 2010

Roger L. Sanders, PE

Chinook Engineering
a division of Inter-Mountain Laboratories, Inc.
555 Absaraka Street
Sheridan, Wyoming 82801 USA
(307) 672-7790
chinook@imlinc.com

Certificate of Accuracy

Transfer Standard Type: Barometric Pressure/Altimeter

Certificate No: B 091009. 02

Transfer standard model: Suunto Escape 203 Electronic Altimeter/Barometer
Serial number: 74605246 IML 1404
submitted by/owner: Inter-Mountain Laboratories, Inc.
Air Science Division
555 Absaraka Street
Sheridan, WY 82801

Was compared to Precision Absolute Reference Barometer:
Model number: 355-AI0900 Serial number: 913930-M1
Certified accuracy of ± 0.007"Hg
NIST traceable to Ruska Deadweight Tester SN 38342/C-85

Date: 09/10/09 Lab temperature 75.7 °F
Lab pressure 663.2 mm Hg

Reference barometer ("Hg)	Transfer Standard ("Hg)	Difference from Reference ("Hg)	Transfer Standard Correction* ("Hg)
24.00	23.95	-0.05	0.05
25.00	24.95	-0.05	0.05
26.10	26.05	-0.05	0.05
26.50	26.45	-0.05	0.05
27.55	27.50	-0.05	0.05

Note:
If no sign is given on the correction, the true pressure is higher than the indicated pressure. If the sign is negative, the true pressure is lower than the indicated pressure.

Transfer Standard adjustments made? YES ☐ NO ☑

Post-calibration measurements:

Reference barometer ("Hg)	Transfer Standard ("Hg)	Difference from Reference ("Hg)	Transfer Standard Correction* ("Hg)

Certified By: Date: September 10, 2009

Roger L. Sanders, PE

Chinook Engineering
a division of Inter-Mountain Laboratories, Inc.
555 Absaraka Street
Sheridan, Wyoming 82801 USA
(307) 672-7790
chinook@imlinc.com

Certificate of Accuracy

Transfer Standard Type: Electronic Hygrometer Certificate No: H 021110. 01

Transfer standard, model/type: Dwyer Series 485 Digital Hygrometer
Serial number: IML 0899 (Previously IML 0936)
submitted by/owner: Inter-Mountain Laboratories, Inc.
Air Science Division
555 Absaraka Street
Sheridan, WY 82801

Was compared to Saturated Salt Solution Standards using ASTM Method E 104 - 02, Standard Practice for Maintaining Constant Relative Humidity by Means of Aqueous Solutions, using Temperature Reference Standard Streamline™ Pro MultiCal™ System Remote Temperature Probe S/N T030301.

Date: 02/10/2010 - 02/11/2010

Lab temperature: 68 - 72 °F
Barometric Pressure: 657 - 659 mmHG
Lab %RH: 35 - 45%

Reference Salt Standard	Reference Temperature °C	Reference Standard (%RH)	Uncertainty	Transfer Standard (%RH)	Difference from Reference (%RH)	Transfer Standard Correction* (%RH)
Lithium Chloride	18.5	12.06	±0.90	7.5	-4.6	4.6
Potassium Acetate						
Magnesium Chloride	18.7	33.13	±0.23	27.6	-5.5	5.5
Magnesium Nitrate	18.9	54.71	±0.24	49.6	-5.1	5.1
Sodium Chloride	19.0	75.50	±0.20	72.2	-3.3	3.3

Temperature Reference Standard (°C)		Transfer Standard (°C)	Difference from Reference (°C)	Transfer Standard Correction* (°C)
18.5		18.3	-0.2	0.2
18.7		18.5	-0.2	0.2
18.9		18.6	-0.3	0.3
19.0		18.7	-0.3	0.3

Certified By: _____ Date: February 11, 2010

Roger L. Sanders, PE

Chinook Engineering
a division of Inter-Mountain Laboratories, Inc.
555 Absaraka Street
Sheridan, Wyoming 82801 USA
(307) 672-7790
chinook@imlinc.com

CERTIFICATE OF CALIBRATION

Test Unit

Model:	18802
Motor SN:	IML 0896
Control Unit SN:	CA 03156
Range:	200 - 15000

Motor RPM	Indicated RPM					
	Clockwise Output Range			Counter Clockwise Output Range		
	Low	High	Average	Low	High	Average
600	599.9	600.1	599.9	600.0	600.0	600.0
1200	1199.9	1200.1	1199.9	1199.3	1200.1	1199.8
2400	2399.8	2399.8	2399.8	2399.8	2399.8	2399.8
4200	4199.7	4199.8	4199.7	4199.7	4199.8	4199.7
6000	5999.7	5999.7	5999.7	5999.7	6000.6	5999.7
8100	8099.6	8099.6	8099.6	8099.6	8099.6	8099.6
9900	9899.5	9899.5	9899.5	9899.5	9899.5	9899.5

The insrument above has been presented for inspection and test as shown
The indicated work was performed using standards traceable to the National Institute of Standards and Technologies (NIST)

Standard SN:	51892014
Technician:	C. Medill
Date:	1/22/2010

THE BRUNTON COMPANY
Certificate Of Calibration

IML 0942

Equipment Owner:
Name: Intec - Mountain Labs

Address: 555 Absarada St.

City, State, Zip: Sheridan Wyo 82801

Calibration traceable to the National Institute of Standards and Technology in accordance with Mil-STD-45662A has been accomplished on the instrument listed below by comparison with standards maintained by The Brunton Co. The accuracy and stability of all standards maintained by The Brunton Co. are traceable to national standards maintained by the National Institute of Standards and Technology in Washington, D.C. and Boulder, CO. Complete record of all work performed is maintained by The Brunton Co. and is available for inspection upon request.

This Unit has been calibrated to Lietz TM10E serial number 30937 traceable to N.B.S. no. 738 227675 this 18 Day of February 20 10

DESCRIPTION: Pocket transit

PURCHASE ORDER: 19845

ORDER NUMBER: 251518

LOT NUMBER: 2081319

MODEL NUMBER: 5004

SERIAL NUMBER: 5080393535

CALIBRATION DATE: 2/18/10

RECALIBRATION DUE DATE: 2/18/11

Signed: Eric Ingalls
QUALITY CONTROL MANAGER

apogee instruments inc.
721 W 1800 N
Logan, UT 84321

Certificate of Calibration

LI-COR PYRANOMETER
MODEL LI-200X

Customer Name	:	Kristy Kola
Serial Number	:	PY54289
Calibration Date	:	17-Apr-2009
Previous Calibration Date	:	12-Aug-2007
Recommended Recalibration Date	:	17-Apr-2011
Calibration Factor	:	**208.60 W m^{-2} per mV**
Output	:	**85.3 µA per 1000 W m^{-2}**
Calibration Factor as Received	:	198.37 W m^{-2} per mV
Output as Received	:	89.7 µA per 1000 W m^{-2}
Resistor Size (Measured)	:	56.2 Ω
Percent Change in Output	:	-4.91%
Percent Change per Year	:	-2.92%

Calibration Procedure

□ Calibrated to a heated and ventilated Kipp & Zonen CM21 pyranometer under sunlight for an integrated daily total in Logan, Utah.

☒ Calibrated to four LI-COR transfer standard pyranometers under Metal Halide lamps. The transfer standards are calibrated to the Kipp & Zonen CM21.

Traceability

The Kipp & Zonen Model CM21 pyranometer is calibrated every two years at NREL in Boulder, CO. NREL reference standards are calibrated every five years to the world meteorological reference in Davos, Switzerland.

Calibration Technician : _____ Date : Apr 17, 2009

Web page: www.apogee-inst.com E-mail: admin@apogee-inst.com Tel: (435)792-4700 Fax: (435)787-8268